

Certificate of Analysis

PREPARED MICROBIOLOGICAL MEDIA



Catalogue Number CMG105-10-10
Product Description Brain Heart Infusion Agar, 10cm Petri Dish, Package of 10
Lot Number 1040
Date of Manufacture 1 November 2021
Expiry Date 30 December 2021

QA Testing:	Result:	Expected:
Characteristics	Pass	Pass
Sterility	Pass	Pass
Performance	Pass	Pass

pH 7.4
Appearance Yellow, Clear
Storage Condition Refrigerate, 2-8°C

Sterility Method Autoclave
Sterility Test Pass
(Absence of growth following 72 hours at 30 - 35°C)

Organisms	Incubation	Results
<i>Staphylococcus aureus</i> ATCC 25923	35 ± 2°C for 18-48 hours	Growth
<i>Escherichia coli</i> ATCC 25922	35 ± 2°C for 18-48 hours	Growth
<i>Aspergillus brasiliensis (niger)</i> ATCC 16404	30 ± 2°C for 18-72 hours	Growth

Released By: Mehdi Kargar

Date: 4 November 2021

All our prepared media products are manufactured at our site in RCFFN, University of Manitoba and tested both at our site and by department of Microbiology, University of Manitoba.

The generation of this certificate confirms all sterilization and performance criteria have been achieved.

NOTE: Expiry Date only valid if packs stored unopened at Ambient Room Temperature not exceeding 25°C.

CBSAlife

155-196 Innovation Drive, Winnipeg, MB, R3T 2N2, Canada

Phone: +1 (204) 269-2255

Email: info@cbsalife.com

Website: <https://cbsalife.com>