

Certificate of Analysis

PREPARED MICROBIOLOGICAL MEDIA



Catalogue Number CMB141-10-10
Product Description Bile Esculin Agar, 10cm Petri Dish, Package of 10
Lot Number 1063
Date of Manufacture 9 March 2022
Expiry Date 23 May 2022

QA Testing:	Result:	Expected:
Characteristics	Pass	Pass
Sterility	Pass	Pass
Performance	Pass	Pass

pH 6.8 @ 25°C
Appearance Dark, tan olive to olive green with a blue tint, trace hazy to hazy
Storage Condition Refrigerate, 2-8°C

Sterility Method Autoclave
Sterility Test Pass
(Absence of growth following 72 hours at 30 - 35°C)

Organisms	Incubation	Results
<i>Enterococcus faecalis</i> ATCC 29212	35 ± 2°C for 42-48 hours	Growth, Blackening

Released By: Mehdi Kargar

Date: 12 March 2022

All our prepared media products are manufactured at our site in RCFFN, University of Manitoba and tested both at our site and by department of Microbiology, University of Manitoba.

The generation of this certificate confirms all sterilization and performance criteria have been achieved.

NOTE: Expiry Date only valid if packs stored unopened at Ambient Room Temperature not exceeding 25°C.

CBSAlife

155-196 Innovation Drive, Winnipeg, MB, R3T 2N2, Canada

Phone: +1 (204) 269-2255

Email: info@cbsalife.com

Website: <https://cbsalife.com>