

Certificate of Analysis

PREPARED MICROBIOLOGICAL MEDIA



Catalogue Number CMB147-10-10
Product Description DNase Test Agar with Toluidine Blue, 10cm Petri Dish, Package of 10
Lot Number 1093
Date of Manufacture 2 May 2022
Expiry Date 16 July 2022

QA Testing:	Result:	Expected:
Characteristics	Pass	Pass
Sterility	Pass	Pass
Performance	Pass	Pass

pH 7.3 @ 25°C
Appearance Medium to Dark, Blue, Hazy
Storage Condition Refrigerate, 2-8°C

Sterility Method Autoclave
Sterility Test Pass
(Absence of growth following 72 hours at 30 - 35°C)

Organisms	Incubation	Results
<i>Enterobacter aerogenes</i> ATCC 13048	35 ± 2°C for 18-24 hours	Growth, DNase -
<i>Klebsiella pneumoniae</i> ATCC 33495	35 ± 2°C for 18-24 hours	Growth, DNase -
<i>Serratia marcescens</i> ATCC 13880	30 ± 2°C for 18-24 hours	Growth, DNase +

Released By: Mehdi Kargar

Date: 5 May 2022

All our prepared media products are manufactured at our site in RCFFN, University of Manitoba and tested both at our site and by department of Microbiology, University of Manitoba.

The generation of this certificate confirms all sterilization and performance criteria have been achieved.

NOTE: Expiry Date only valid if packs stored unopened at Ambient Room Temperature not exceeding 25°C.

CBSA life

155-196 Innovation Drive, Winnipeg, MB, R3T 2N2, Canada

Phone: +1 (204) 269-2255

Email: info@cbsalife.com

Website: <https://cbsalife.com>