

## Certificate of Analysis

### PREPARED MICROBIOLOGICAL MEDIA



Catalogue Number CMS158-10-10  
Product Description Listeria Selective Agar, 10cm Petri Dish, Package of 10.  
Lot Number 1010  
Date of Manufacture 6 September 2021  
Expiry Date 20 November 2021

<b>QA Testing:</b>	<b>Result:</b>	<b>Expected:</b>
<b>Characteristics</b>	Pass	Pass
<b>Sterility</b>	Pass	Pass
<b>Performance</b>	Pass	Pass

pH 7.3 @ 25°C  
Appearance Light to Medium Amber, Slightly Opalescent  
Storage Condition Refrigerate, 2-8°C

Sterility Method Autoclave  
Sterility Test Pass  
(Absence of growth following 72 hours at 30 - 35°C)

<b>Organisms</b>	<b>Incubation</b>	<b>Results</b>
<i>Escherichia coli</i> ATCC 25922	35 ± 2°C for 18-48 hours	Complete Inhibition
<i>Listeria monocytogenes</i> ATCC 19114	35 ± 2°C for 18-48 hours	Growth, Black Colonies

Released By: Mehdi Kargar

Date: 9 September 2021

All our prepared media products are manufactured at our site in RCFFN, University of Manitoba and tested both at our site and by department of Microbiology, University of Manitoba.

The generation of this certificate confirms all sterilization and performance criteria have been achieved.

NOTE: Expiry Date only valid if packs stored unopened at Ambient Room Temperature not exceeding 25°C.

**CBSA**life

155-196 Innovation Drive, Winnipeg, MB, R3T 2N2, Canada

Phone: +1 (204) 269-2255

Email: [info@cbsalife.com](mailto:info@cbsalife.com)

Website: <https://cbsalife.com>