

Certificate of Analysis

PREPARED MICROBIOLOGICAL MEDIA



Catalogue Number CMG156-10-10
Product Description Luria-Bertani (LB) Broth, 10ml Tube, Package of 10
Lot Number 1089
Date of Manufacture 22 April 2022
Expiry Date 21 June 2022

QA Testing:	Result:	Expected:
Characteristics	Pass	Pass
Sterility	Pass	Pass
Performance	Pass	Pass

pH 7.0 @ 25°C
Appearance Very light Amber, Clear to very Slightly Opalescent.
Storage Condition Refrigerate, 2-8°C

Sterility Method Autoclave
Sterility Test Pass
(Absence of growth following 72 hours at 30 - 35°C)

Organisms	Incubation	Results
<i>Escherichia coli</i> ATCC 33526	35 ± 2°C for 18-48 hours	Growth

Released By: Mehdi Kargar

Date: 25 April 2022

All our prepared media products are manufactured at our site in RCFFN, University of Manitoba and tested both at our site and by department of Microbiology, University of Manitoba.

The generation of this certificate confirms all sterilization and performance criteria have been achieved.

NOTE: Expiry Date only valid if packs stored unopened at Ambient Room Temperature not exceeding 25°C.

CBSAlife

155-196 Innovation Drive, Winnipeg, MB, R3T 2N2, Canada

Phone: +1 (204) 269-2255

Email: info@cbsalife.com

Website: <https://cbsalife.com>