

Certificate of Analysis

PREPARED MICROBIOLOGICAL MEDIA



Catalogue Number CMS159-6-10
Product Description m-Endo Agar, 6cm Petri Dish, Package of 10
Lot Number 1080
Date of Manufacture 02 May 2022
Expiry Date 16 July 2022

QA Testing:	Result:	Expected:
Characteristics	Pass	Pass
Sterility	Pass	Pass
Performance	Pass	Pass

pH 7.2 @ 25°C
Appearance Rose colored, slightly opalescent, with precipitate.
Storage Condition Refrigerate, 2-8°C

Sterility Method Autoclave
Sterility Test Pass
(Absence of growth following 72 hours at 30 - 35°C)

Organisms	Incubation	Results
<i>Escherichia coli</i> ATCC 25922	35 ± 2°C for 20-24 hours	Growth, Red with sheen colonies
<i>Salmonella enterica</i> ATCC 14028	35 ± 2°C for 20-24 hours	Growth, Pink colonies
<i>Staphylococcus aureus</i> ATCC 25923	35 ± 2°C for 20-24 hours	Inhibited

Released By: Mehdi Kargar

Date: 5 May 2022

All our prepared media products are manufactured at our site in RCFFN, University of Manitoba and tested both at our site and by department of Microbiology, University of Manitoba.

The generation of this certificate confirms all sterilization and performance criteria have been achieved.

NOTE: Expiry Date only valid if packs stored unopened at Ambient Room Temperature not exceeding 25°C.

CBSA life

155-196 Innovation Drive, Winnipeg, MB, R3T 2N2, Canada

Phone: +1 (204) 269-2255

Email: info@cbsalife.com

Website: <https://cbsalife.com>