

Certificate of Analysis

PREPARED MICROBIOLOGICAL MEDIA



Catalogue Number CMS173-15-10
Product Description m-FC Agar with BCIG, 15cm Petri Dish, Package of 10
Lot Number 1102
Date of Manufacture 24 May 2022
Expiry Date 22 August 2022

QA Testing:	Result:	Expected:
Characteristics	Pass	Pass
Sterility	Pass	Pass
Performance	Pass	Pass

pH 7.2 @ 25°C
Appearance Light amber
Storage Condition Refrigerate, 2-8°C

Sterility Method Autoclave
Sterility Test Pass
(Absence of growth following 72 hours at 30 - 35°C)

Organisms	Incubation	Results
<i>Escherichia coli</i> ATCC 25922	35 ± 2°C for 18-24 hours	Growth, Blue colonies
<i>Salmonella typhimurium</i> ATCC 14028	35 ± 2°C for 18-24 hours	Growth, Clear colonies
<i>Klebsiella pneumoniae</i> ATCC 13883	35 ± 2°C for 18-24 hours	Growth, Cream colonies
<i>Enterococcus faecalis</i> ATCC 29212	35 ± 2°C for 18-24 hours	Inhibited

Released By: Mehdi Kargar

Date: 27 May 2022

All our prepared media products are manufactured at our site in RCFFN, University of Manitoba and tested both at our site and by department of Microbiology, University of Manitoba.

The generation of this certificate confirms all sterilization and performance criteria have been achieved.

NOTE: Expiry Date only valid if packs stored unopened at Ambient Room Temperature not exceeding 25°C. Keep away from direct light.



155-196 Innovation Drive, Winnipeg, MB, R3T 2N2, Canada

Phone: +1 (204) 269-2255

Email: info@cbsalife.com

Website: <https://cbsalife.com>