

## Certificate of Analysis

### PREPARED MICROBIOLOGICAL MEDIA



Catalogue Number CMS131-10-10  
Product Description Mannitol Salt Agar, 10cm Petri Dish, Package of 10.  
Lot Number 1015  
Date of Manufacture 26 November 2021  
Expiry Date 9 February 2022

QA Testing:	Result:	Expected:
Characteristics	Pass	Pass
Sterility	Pass	Pass
Performance	Pass	Pass

pH 7.4 @ 25°C  
Appearance Red, Clear  
Storage Condition Refrigerate, 2-8°C

Sterility Method Autoclave  
Sterility Test Pass  
(Absence of growth following 72 hours at 30 - 35°C)

Organisms	Incubation	Results
<i>Staphylococcus aureus</i> ATCC 25923	35 ± 2°C for 24-48 hours	Growth, Yellow
<i>Staphylococcus epidermidis</i> ATCC 12228	35 ± 2°C for 24-48 hours	Growth, Red

Released By: Mehdi Kargar

Date: 1 December 2021

All our prepared media products are manufactured at our site in RCFFN, University of Manitoba and tested both at our site and by department of Microbiology, University of Manitoba.

The generation of this certificate confirms all sterilization and performance criteria have been achieved.

NOTE: Expiry Date only valid if packs stored unopened at Ambient Room Temperature not exceeding 25°C.

**CBSA**life

155-196 Innovation Drive, Winnipeg, MB, R3T 2N2, Canada

Phone: +1 (204) 269-2255

Email: [info@cbsalife.com](mailto:info@cbsalife.com)

Website: <https://cbsalife.com>