

Certificate of Analysis

PREPARED MICROBIOLOGICAL MEDIA



Catalogue Number CMG102-10-10
Product Description Nutrient Broth, 10ml/tube, Package of 10
Lot Number 1081
Date of Manufacture 22 April 2022
Expiry Date 21 June 2022

QA Testing:	Result:	Expected:
Characteristics	Pass	Pass
Sterility	Pass	Pass
Performance	Pass	Pass

pH 6.8 @ 25°C
Appearance Light Yellow, Clear
Storage Condition Refrigerate, 2-8°C

Sterility Method Autoclave
Sterility Test Pass
(Absence of growth following 72 hours at 30 - 35°C)

Organisms	Incubation	Results
<i>Staphylococcus aureus</i> ATCC 25923	35 ± 2°C for 18-48 hours	Good Growth
<i>Escherichia coli</i> ATCC 25922	35 ± 2°C for 18-48 hours	Good Growth

Released By: Mehdi Kargar

Date: 25 April 2022

All our prepared media products are manufactured at our site in RCFFN, University of Manitoba and tested both at our site and by department of Microbiology, University of Manitoba.

The generation of this certificate confirms all sterilization and performance criteria have been achieved.

NOTE: Expiry Date only valid if packs stored unopened at Ambient Room Temperature not exceeding 25°C.

CBSAlife

155-196 Innovation Drive, Winnipeg, MB, R3T 2N2, Canada

Phone: +1 (204) 269-2255

Email: info@cbsalife.com

Website: <https://cbsalife.com>