

## Certificate of Analysis

### PREPARED MICROBIOLOGICAL MEDIA



|                     |                                                       |
|---------------------|-------------------------------------------------------|
| Catalogue Number    | CMS122-10-10                                          |
| Product Description | Sabouraud 4% Dextrose Broth, 10ml Tube, Package of 10 |
| Lot Number          | 1090                                                  |
| Date of Manufacture | 2 May 2022                                            |
| Expiry Date         | 1 July 2022                                           |

| QA Testing:     | Result: | Expected: |
|-----------------|---------|-----------|
| Characteristics | Pass    | Pass      |
| Sterility       | Pass    | Pass      |
| Performance     | Pass    | Pass      |

|                   |                    |
|-------------------|--------------------|
| pH                | 5.6 @ 25°C         |
| Appearance        | Light Amber, Clear |
| Storage Condition | Refrigerate, 2-8°C |

|                                                     |           |
|-----------------------------------------------------|-----------|
| Sterility Method                                    | Autoclave |
| Sterility Test                                      | Pass      |
| (Absence of growth following 72 hours at 30 - 35°C) |           |

| Organisms                                          | Incubation                | Results     |
|----------------------------------------------------|---------------------------|-------------|
| <i>Candida albicans</i> ATCC 10231                 | 30 ± 2°C for up to 7 days | Good Growth |
| <i>Saccharomyces cerevisiae</i> ATCC 9763          | 30 ± 2°C for up to 7 days | Good Growth |
| <i>Aspergillus brasiliensis (niger)</i> ATCC 16404 | 30 ± 2°C for up to 7 days | Good Growth |

Released By: Mehdi Kargar

Date: 5 May 2022

All our prepared media products are manufactured at our site in RCFFN, University of Manitoba and tested both at our site and by department of Microbiology, University of Manitoba.

The generation of this certificate confirms all sterilization and performance criteria have been achieved.

NOTE: Expiry Date only valid if packs stored unopened at Ambient Room Temperature not exceeding 25°C.

**CBSA**life

155-196 Innovation Drive, Winnipeg, MB, R3T 2N2, Canada

Phone: +1 (204) 269-2255

Email: [info@cbsalife.com](mailto:info@cbsalife.com)

Website: <https://cbsalife.com>