

## Certificate of Analysis

### PREPARED MICROBIOLOGICAL MEDIA



Catalogue Number CMD170-10-10  
Product Description Sheep Blood Tryptic Soy Agar with 0.1% Esculin, 10cm Petri Dish, Package of 10  
Lot Number 1065  
Date of Manufacture 12 April 2022  
Expiry Date 11 July 2022

QA Testing:	Result:	Expected:
Characteristics	Pass	Pass
Sterility	Pass	Pass
Performance	Pass	Pass

pH 7.3  
Appearance Red, Opaque  
Storage Condition Refrigerate, 2-8°C

Sterility Method Autoclave  
Sterility Test Pass  
(Absence of growth following 72 hours at 30 - 35°C)

Organisms	Incubation	Results
<i>Enterococcus faecalis</i> ATCC 29212	35 ± 2°C for 18-48 hours	Growth, Esculin Positive
<i>Staphylococcus aureus</i> ATCC 25923	35 ± 2°C for 18-48 hours	Growth, Esculin Positive (light)
<i>Streptococcus agalactiae</i> ATCC 12386	35 ± 2°C for 18-48 hours	Growth, with No darkening around colony
<i>Streptococcus pyogenes</i> ATCC 19619	35 ± 2°C for 18-48 hours	Partial Growth, Esculin Negative
<i>Streptococcus pneumoniae</i> ATCC 6305	35 ± 2°C for 18-48 hours	Partial Growth, Esculin Negative

Released By: Mehdi Kargar

Date: 18 April 2022

All our prepared media products are manufactured at our site in RCFFN, University of Manitoba and tested both at our site and by department of Microbiology, University of Manitoba.

The generation of this certificate confirms all sterilization and performance criteria have been achieved.

NOTE: Expiry Date only valid if packs stored unopened at Ambient Room Temperature not exceeding 25°C.



155-196 Innovation Drive, Winnipeg, MB, R3T 2N2, Canada

Phone: +1 (204) 269-2255

Email: [info@cbsalife.com](mailto:info@cbsalife.com)

Website: <https://cbsalife.com>