Certificate of Analysis

PREPARED MICROBIOLOGICAL MEDIA



Catalogue Number CMB138-7-10

Product Description Simmons Citrate Agar, Slant Tube, Package of 10

Lot Number 1032

Date of Manufacture 15 October 2021 Expiry Date 14 January 2021

QA Testing:	Result:	Expected:
Characteristics	Pass	Pass
Sterility	Pass	Pass
Performance	Pass	Pass

pH 6.8 @ 25°C

Appearance Dark Green, Slightly Opalescent

Storage Condition Refrigerate, 2-8°C

Sterility Method Autoclave Sterility Test Pass

(Absence of growth following 72 hours at 30 - 35°C)

Organisms	Incubation	Results
Escherichia coli ATCC 25922	35 ± 2 °C for 18-24 hours	Growth, Alkaline (Blue)
Klebsiella pneumoniae ATCC 12022	35 ± 2 °C for 18-24 hours	Growth, Alkaline (Blue)

Released By: Mehdi Kargar

Date: 19 October 2021 Mehds

All our prepared media products are manufactured at our site in RCFFN, University of Manitoba and tested both at our site and by department of Microbiology, University of Manitoba.

The generation of this certificate confirms all sterilization and performance criteria have been achieved.

NOTE: Expiry Date only valid if packs stored unopened at Ambient Room Temperature not exceeding 25°C.

CBSA life

155-196 Innovation Drive, Winnipeg, MB, R3T 2N2, Canada

Phone: +1 (204) 269-2255 Email: info@cbsalife.com Website: https://cbsalife.com