

Certificate of Analysis

PREPARED MICROBIOLOGICAL MEDIA



Catalogue Number	CMG103-10-10
Product Description	Tryptic Soy Agar (TSA), 10cm Petri Dish, Package of 10
Lot Number	1022
Date of Manufacture	1 October 2021
Expiry Date	14 December 2021

QA Testing:	Result:	Expected:
Characteristics	Pass	Pass
Sterility	Pass	Pass
Performance	Pass	Pass

pH	7.4 @ 25°C
Appearance	Amber, Clear
Storage Condition	Refrigerate, 2-8°C

Sterility Method	Autoclave
Sterility Test	Pass
(Absence of growth following 72 hours at 30 - 35°C)	

Organisms	Incubation	Results
<i>Staphylococcus aureus</i> ATCC 25923	35 ± 2°C for 18-48 hours	Growth
<i>Escherichia coli</i> ATCC 25922	35 ± 2°C for 18-48 hours	Growth

Released By: Mehdi Kargar

Date: 4 October 2021

All our prepared media products are manufactured at our site in RCFFN, University of Manitoba and tested both at our site and by department of Microbiology, University of Manitoba.

The generation of this certificate confirms all sterilization and performance criteria have been achieved.

NOTE: Expiry Date only valid if packs stored unopened at Ambient Room Temperature not exceeding 25°C.

CBSA life

155-196 Innovation Drive, Winnipeg, MB, R3T 2N2, Canada

Phone: +1 (204) 269-2255

Email: info@cbsalife.com

Website: <https://cbsalife.com>