

Certificate of Analysis

PREPARED MICROBIOLOGICAL MEDIA



Catalogue Number CMS129-10-10
Product Description Xylose Lysine Deoxycholate (XLD) Agar, 10cm Petri Dish,
Package of 10.
Lot Number 1057
Date of Manufacture 26 November 2021
Expiry Date 9 February 2022

QA Testing:	Result:	Expected:
Characteristics	Pass	Pass
Sterility	Pass	Pass
Performance	Pass	Pass

pH 7.4 @ 25°C
Appearance Red, Clear
Storage Condition Refrigerate, 2-8°C

Sterility Method Autoclave
Sterility Test Pass
(Absence of growth following 72 hours at 30 - 35°C)

Organisms	Incubation	Results
<i>Escherichia coli</i> ATCC 25922	35 ± 2°C for 18-24 hours	Growth, Yellow
<i>Shigella flexneri</i> ATCC 12022	35 ± 2°C for 18-24 hours	Growth, Red

Released By: Mehdi Kargar

Date: 1 December 2021

All our prepared media products are manufactured at our site in RCFFN, University of Manitoba and tested both at our site and by department of Microbiology, University of Manitoba.

The generation of this certificate confirms all sterilization and performance criteria have been achieved.

NOTE: Expiry Date only valid if packs stored unopened at Ambient Room Temperature not exceeding 25°C.

CBSAlife

155-196 Innovation Drive, Winnipeg, MB, R3T 2N2, Canada

Phone: +1 (204) 269-2255

Email: info@cbsalife.com

Website: <https://cbsalife.com>